

## Manufacturer's Certificate of Validation

Certificate No	Date	Page
VC182300062	May 12 , 2018	1 of 1

SENCIUS TECHNOLOGIES certifies that the products mentioned on the following page(s) have been thoroughly tested, validated and met performance accuracy specifications over the stated ranges.

### Reference Instrumentation

Device	Model	Calibration Date	Accuracy/Volatility	Serial Number
Constant Temperature Trough of Low Temperature	HC-2010	March 25,2017	±0.05°C	JYDQ-135
Low-Temperature checking Equipment	ATL60	March 25,2017	±0.015°C	JYDQ-136

All reference instruments are calibrated by Nanjing Test Technology Co., LTD, a CNAS (China National Accreditation Service) accredited laboratory, and traceable to NIST according to Multi-Recognition Agreement

### Environment Conditions

<b>Air Temperature</b>
25.0°C
<b>Relative Humidity</b>
53.0%

### Validation Information

Validation Points	Tolerance	Result
+40°C	±0.5°C	PASS
0°C	±0.5°C	PASS
-20°C	±0.5°C	PASS

### Product Information

Product Name	Serial Number	Production Date	Expiry Date	Quantity
lite	EF2182302441 - EF2182302480	02/2018	02/2020	40

This certificate was automatically generated and valid for 2(two) years from date of production.

Approved by:

*Leo Zhang*

